

RESULTS FORM TO BE FILLED IN BY THE PARTICIPATING LABORATORY

PT trial Title	Detection of Equine Infectious Anemia (EIA) antibodies using ELISA		
Identification of the participating laboratory			
Laboratory name	CLINICA EQUINA srl		
Full name of the contact person	Teótimo Becú		
Address	Alem 1698 - Capitan Sarmiento - CP 2752 - Buenos Aires, ARGENTINA		
☎ 0054 2478 481489	Mail	t-becu@redsarmiento.com.ar	Fax 0054 2478 481658
Schedule of the PT trial			
Date of samples analysis	22-6-2018		
Results due by	Teótimo Becú		
Methods or modus operandi used	CELISA read at 620 nm.		
ELISA Kit used IDEXX CELISA			
Batch numbers EP 301			
Expiration date 05-01-2019			
Miscellaneous information	Cut off: 0,481 . Samples with that OD or less, mustbe considered POSITIVES.		
This document has one attached annex			
Signature of the contact person for this PT trial			

Deadline for sending results to : **July the 6th, 2018 (6 p.m).**

Full name	Anses Laboratory for equine diseases		
Name	Delphine GAUDAIRE	Unit	VIPARE
Address	RD675		
Zip code	14430	City	GOUSTRANVILLE
☎ + 33 (0)2 31 79 22 76	Mail	delphine.gaudaire@anses.fr	Fax + 33 (0)2 31 39 21 37


Dr. TEÓTIMO BECÚ
 Médico Veterinario - M.P. 2007
 CLINICA EQUINA S.R.L.
 MAJADON DE LA LA 0124

RESULTS FORM TO BE FILLED IN BY THE PARTICIPATING LABORATORY


NOTE : Each vial must be reconstituted with 500 µl of sterile water. Shake thoroughly (vortex) until complete dissolution pellet. Incubate at room temperature (18-25°C) for 30 minutes while shaking the vial time to time. After reconstitution sera samples have to be analyzed using ELISA following manufacturer's instructions.

Sera sample numbers are listed in the chart 1 below.

Could you, please, fill out the chart « EIA-ELISA results» (positive/negative)

If you have any questions regarding methods, please feel free to contact the PT coordinator at Delphine.gaudaire@anses.fr.

Sample codification	ELISA Results	OD 620 nm
1-18-AN-8	NEGATIVE	0,715
2-18-AN-32	POSITIVE	0,11
3-18-AN-84	NEGATIVE	0,628
4-18-AN-133	POSITIVE	0,208
5-18-AN-192	POSITIVE	0,09
6-18-AN-352	POSITIVE	0,094
7-18-AN-356	POSITIVE	0,094
8-18-AN-468	NEGATIVE	0,664
9-18-AN-520	POSITIVE	0,101
10-18-AN-526	POSITIVE	0,122
11-18-AN-537	POSITIVE	0,112
12-18-AN-545	POSITIVE	0,262
13-18-AN-567	NEGATIVE	0,842
14-18-AN-605	NEGATIVE	0,785
15-18-AN-617	NEGATIVE	0,825
16-18-AN-623	POSITIVE	0,098
17-18-AN-676	POSITIVE	0,101
18-18-AN-718	NEGATIVE	0,77
19-18-AN-760	POSITIVE	0,081
20-18-AN-763	POSITIVE	0,252
21-18-AN-827	POSITIVE	0,121
22-18-AN-948	NEGATIVE	0,775
23-18-AN-975	POSITIVE	0,073
24-18-AN-987	POSITIVE	0,195
	Ref POSITIVE	0,481
	Ref NEGATIVE	0,664


Dr. TOTO MO NECU
 VETERINAIRE - M.D. OMS
 CHIRURGIE VETERINAIRE S.A.L.
 LABORATOIRE DE SECURITE LA 0104

Once filled out, please return this document duly signed by e-mail, mail or fax.

1.

DOCUMENTO EXTI
CONTROLADY

PRUEBA DE PROFICIENCIA CELISA AIE -ANSES- FRANCIA

Teotimo Becu – cutt off 0,566

MUESTRAS	SIEMBRA 22/06/2018	
	RESULTADO	Lectura OD 620 nm
1-18-AN-8	NEGATIVE	0,715
2-18-AN-32	POSITIVE	0,11
3-18-AN-84	NEGATIVE	0,628
4-18-AN-133	POSITIVE	0,208
5-18-AN-192	POSITIVE	0,090
6-18-AN-352	POSITIVE	0,094
7-18-AN-356	POSITIVE	0,094
8-18-AN-468	NEGATIVE	0,664
9-18-AN-520	POSITIVE	0,101
10-18-AN-526	POSITIVE	0,122
11-18-AN-537	POSITIVE	0,112
12-18-AN-545	POSITIVE	0,262
13-18-AN-567	NEGATIVE	0,842
14-18-AN-605	NEGATIVE	0,785
15-18-AN-617	NEGATIVE	0,825
16-18-AN-623	POSITIVE	0,098
17-18-AN-676	POSITIVE	0,101
18-18-AN-718	NEGATIVE	0,77
19-18-AN-760	POSITIVE	0,081
20-18-AN-763	POSITIVE	0,252
21-18-AN-827	POSITIVE	0,121
22-18-AN-948	NEGATIVE	0,775
23-18-AN-975	POSITIVE	0,073
24-18-AN-987	POSITIVE	0,195


Dr. TEOTIMO BECÚ
MEDICO VETERINARIO - M.P. 889
CLÍNICA SOLIDA S.A.S.
HABITACION DE LA 1004

MUESTRAS : ANSES PROFICIENCY Interpretación: **NEGATIVO**, el OD debe ser mayor al **POSITIVO** de referencia
Cutt off : .566 **POSITIVO**, el OD es igual o menor al **POSITIVO** de referencia

	1	2	3	4	5	6	7	8	9	10	11	12
A												
B	1- Neg 0,715	2- POS 0,11	3- Neg 0,628	4- POS 0,208	5- POS 0,09	6- POS 0,094	7- POS 0,094	8- Neg 0,664	9- POS 0,101	10- POS 0,122	11- POS 0,112	12- POS 0,262
C	13- Neg 0,842	14- Neg 0,785	15- Neg 0,825	16- POS 0,098	17- POS 0,101	Posit ref 0,481	18- Neg 0,77	19- POS 0,081	20- POS 0,252	21- POS 0,121	22- Neg 0,775	23- POS 0,073
D	24- POS 0,195	Posit ref 0,566	Negat ref 1,001									
E												
F												
G												
H												

Dr. EDUARDO BECÚ
 Médico Veterinario - M.P. 1987
 CELISA AIE IDEXX S.A.L.
 Montevideo, Uruguay LA 0134

De: GAUDAIRE Delphine [mailto:Delphine.GAUDAIRE@anses.fr]

Enviado el: martes, 04 de septiembre de 2018 06:00 a.m.

Para: 'Teotimo Becú - Clínica Equina'

Asunto: EIA ELISA PT 2018 results

Dear Teotimo,

You will find below the results that your laboratory obtained to the EIA ELISA PT organized by the EU-RL.

The following criteria were applied to determine whether PT results were satisfactory:

- All negative sera should be found as such
- All positive sera should be found as such

	LAB CODE 18-AN RESULTS	EXPECTED RESULTS
18-AN-567	NEGATIVE	Negative
18-AN-617	NEGATIVE	Negative
18-AN-84	NEGATIVE	Negative
18-AN-948	NEGATIVE	Negative
18-AN-718	NEGATIVE	Negative
18-AN-8	NEGATIVE	Negative
18-AN-468	NEGATIVE	Negative
18-AN-605	NEGATIVE	Negative
18-AN-352	POSITIVE	Positive
18-AN-537	POSITIVE	Positive
18-AN-520	POSITIVE	Positive
18-AN-623	POSITIVE	Positive
18-AN-676	POSITIVE	Positive
18-AN-32	POSITIVE	Positive
18-AN-975	POSITIVE	Positive
18-AN-763	POSITIVE	Positive
18-AN-133	POSITIVE	Positive
18-AN-545	POSITIVE	Positive
18-AN-760	POSITIVE	Positive
18-AN-192	POSITIVE	Positive
18-AN-356	POSITIVE	Positive
18-AN-827	POSITIVE	Positive
18-AN-987	POSITIVE	Positive
18-AN-526	POSITIVE	Positive

Given criteria, your lab obtained satisfactory results to this PT.

PT results will be presented and discussed during the workshop on Wednesday September 12th, 2018.

Best regards,

Delphine GAUDAIRE